

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	27-08-2025 15:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	27-08-2025 15:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Chemicals And Fertilizers
विभाग का नाम/Department Name	Department Of Fertilizers
संगठन का नाम/Organisation Name	Joint Venture Of Cil, Ntpc, IoCl, Fcil & Hfcl
कार्यालय का नाम/Office Name	Hindustan Urvarak And Rasayan Ltd
कुल मात्रा/Total Quantity	2
वस्तु श्रेणी /Item Category	Gas Chromatography (Q3)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	16 Lakh (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Exemption for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Exemption for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Bidder Turnover,Certificate (Requested in ATC),Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय-सीमा बढ़ाने के लिए आवश्यक न्यूनतम सहभागी विक्रेताओं की संख्या। / Minimum number of bids required to disable automatic bid extension	3

बिड विवरण/Bid Details	
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Gas Chromatography
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	3 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Item wise evaluation/
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
Schedule 1 ईएमडी राशि/EMD Amount (In INR)	106676
Schedule 2 ईएमडी राशि/EMD Amount (In INR)	83352

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	41

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

(c).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Manager E&QC

Hindustan Urvarak And Rasayan Ltd, Barauni Unit, Urvarak Nagar, Barauni, Begusarai, Bihar - 851115
(Sh, Srinu Pitta)

विभाजन/Splitting

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	Yes
--	-----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
---	-----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM_No.1_4_2021_PPD_dated_18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

3. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of

purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

5. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- If number of technically qualified bidders are only 2 or 3.
- If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

मूल्यांकन विधि(मदवार मूल्यांकन विधि) / **Evaluation Method** (Item Wise Evaluation Method)

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

मूल्यांकन अनुसूचियां / Evaluation Schedules	वस्तु/श्रेणी / Item/Category	मात्रा / Quantity
Schedule 1	Gas Chromatography	1
Schedule 2	Gas Chromatography	1

Gas Chromatography (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ / **Technical Specifications**

[* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GAS CHROMATOGRAPH (GC) CONFIGURATION	"GC Detectors inclusive in the scope of supply (Select Only those Detectors, which are part of supply)"	Flame Ionization Detector (FID), Thermal Conductivity Detector (TCD or Hot Wire Detector)
	"Number of FID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	1

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	"Number of TCD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	1
	"Number of ECD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of NPD/FTD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of FPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of PID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of PFPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	Number of Inlets GC can Simultaneously Support	1, 2, 3
	Number of Detectors (Ports) GC can Simultaneously Support	1, 2, 3, 4
	Number of Capillary Columns provided in GC - inclusive in the scope of Supply	0, 1, 2
	Type of Capillary Column	Polar, Non-Polar, Semi-Polar, NA
	"GC Capillary Column Length (Hint:-Select '0' if Not Applicable)"	0, 30, 60, 100 (meter)
	"Number of Packed Columns provided in GC - inclusive in the scope of Supply (Hint:-Select '0' if Not Applicable)"	0, 1, 2
	"GC Packed Column Length (Hint:-Select '0' if Not Applicable)"	0, 3, 6, 9, 12 (foot)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PERFORMANCE	"GC Detector Minimum Detection Level (MDL) "	< 1.4 pg C/s for FID, < 1.8 pg C/s for FID, < 3.0 pg C/s for FID < 400 pg/ml for TCD, < = 1 microgram/ml for TCD, < = 10 microgram/ml for TCD, < 4.5 fg /sec for ECD, < 5.5 fg /sec for ECD, <= 100 fg/sec for ECD, < 0.1pg N /sec for NPD/FTD, < 0.03 pg P /sec for NPD/FTD, <0.5 pg P/s for NPD/FTD, < 45 fg P /sec for FPD, < 100 fg P /sec for FPD, <2.5 pg S /sec for FPD, <2.5 pg S /sec for FPD, <5 pg S /sec for FPD, <10pg S/s for FPD
COLUMN OVEN (GAS CHROMATOGRAPH - GC)	Maximum Operating Temperature (GC Oven)	upto 400 deg. C, upto 450 deg. C (2) Or higher
	Number of Oven Ramps / Plateaus for Temperature programming	6/7 or more, 9/10 or more, 20/21 or more
	Temperature Ramp Rate - Maximum	120 Deg. C/minute or better, 50 Deg. C/minute or better
PNEUMATIC CONTROL	Number of Injection Ports	1, 2
	Maximum Injection Port heating Temperature	upto 400 deg. C (1), upto 450 deg. C (2) Or higher
	Pressure Range	upto100 psi (1), upto 120 psi (2) Or higher
	Pressure set point Resolution (±)(psi)	0.001, 0.01, 0.1 Or lower
	Type of Sample Injection Port	Split, Splitless, split / Splitless, Purged packed Inlet
	Split Ratio	6000:1, 7000:1, 7500:1, NA
AUTO INJECTOR / AUTO SAMPLER	Compatible Liquid Auto Sampler - inclusive in the scope of supply	Yes, No
	Capacity of the Liquid Auto Sampler (Number of Vials)	0, 15, 100, 150 Or higher
	Number of Injection Ports in the Liquid Auto Sampler	0, 1, 2
	Number of Channels in the Liquid Auto Sampler	0, 2
	Minimum Volume of the injected liquid sample	0 - 1 (microliter)
	Maximum Volume of the injected liquid sample	0 - 100 (microliter)
	"Liquid Auto Sampler Features (Select applicable features Only)"	Fully automated Liquid Auto Sampler, Injects with variable speed & varying syringe sizes, Operates through Software (Software inclusive in the scope of supply), NA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
HEAD SPACE SAMPLER	Compatible Head Space Sampler- inclusive in the scope of supply	Yes, No
	Capacity of the Head Space Sampler Carousel (number of Vials)	12 or better, NA
	Head Space Sampler Temperature Range	Upto 200 deg. C or better, NA
PYROLYZER SYSTEM	Compatible Pyrolyzer System for pyrolysis of solid and liquid samples- inclusive in the scope of supply	Yes, No
SALIENT FEATURES OF GAS CHROMATOGRAPH (GC)	"Salient Features of GC (Select applicable features Only)"	Automatic leak testing/detection of Pneumatics, Direct setting of split flow rates and ratios, Flow sensor for control & storage of split ratio, Split and Septum Purge gases, Able to direct inject onto packed and wide-bore capillary columns, Oven power turns-off automatically when the lid/door is opened, Over Heat Protection, Display Unit with built-in Key Board, Controlled by External PC System, NA
SOFTWARE	"Software Features (Select Applicable Features only)"	Compatible with Windows 7 or higher version operating system, Seamless integration & control of all the GC parameters, Interface with Multiple GCs through the comprehensive real-time monitoring and diagnostic software, Real time notification via advisories and indicators, Counters, Chromatographic attributes, Instrument diagnostics, Access to maintenance information, Access to maintenance, run and event logs, Retention Time - Locking module for analyzing target compounds in complex matrices
GENERIC	Display	Through PC (PC is not part of supply), Built-in Display
	Connectivity interface	USB, WIFI, ETHERNET, Bluetooth, RS 232, RJ-45
	Provision to split the sample from single injection to two different detectors and get results for two detectors simultaneously	Yes, No
ACESSORIES	Gas Supply inclusive in the scope of supply	Nitrogen, Hydrogen, Argon, Helium, Zero Air, Not offered
	Capacity of Gas cylinder	47 liters water capacity, Not offered
	Number of Cylinders of each type supplied - inclusive in the scope of supply	0, 1, 2
	Gas cylinders certified by PESO	Yes, No for Gas Cylinder Not Offered
	Regulator - inclusive in the scope of supply	Dual Stage Regulator, NA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	"Other Accessories inclusive in the scope of supply (Select only those Accessories which are part of supply)"	Gas purification panel for all Gases with moisture trap, Hydrocarbon trap & Oxygen trap, Syringe for manual injection x 05 nos, Auto Sampler Vials x 1000 vials with screw cap, Graphite ferrules x 10 nos, Vessel ferrules x 10 nos, Septa for injector x 100 nos, Inlet liner for splitless x 4 nos, Inlet liner for split x 4 nos, O-rings x 10 nos. of each type, Column Cutter x 2 nos, Gas tube Cutter x 2 nos, Tool Kit x 1 no., User / Technical / Maintenance Manual, Standard Samples - NIST / IPL Traceable, NA
ENVIRONMENTAL	Minimum Operating Temperature	0 - 10 (degree Celsius)
	Maximum Operating Temperature	30 - 65 (degree Celsius)
	Operating Humidity (RH) (%) at 40 degree C	Up to 95 (percent)
INSTALLATION & COMMISSIONING / WARRANTY / AVAILABILITY OF SPARES / TRAINING	Warranty on equipment	3.0 (year)
	Number of days onsite training provided	3, 5, 7, 10 (day)
CERTIFICATION / TEST REPORT	"Availability of UL/CE Certificate as per IEC 61010- 1 latest - Safety requirements for electrical equipment for measurement, control, and laboratory use General requirements"	Yes, No
	Conformance to EMC/EMI as per IEC 61326-1 latest - Electrical equipment for measurement, control and laboratory use	Yes, No
	Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification	Yes, No

Additional Specification Parameters - Gas Chromatography (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
GC (TCD+FID) : The Technical Specifications strictly must be as per the Attached ATC Document only.	GC (TCD+FID) : The Technical Specifications strictly must be as per the Attached ATC Document only.

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Kumar	851115,HURL BARAUNI PROJECT, BARAUNI URVARAK NAGAR, BEGUSARAI, Begusarai, Bihar, 851115	1	56

Gas Chromatography (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GAS CHROMATOGRAPH (GC) CONFIGURATION	"GC Detectors inclusive in the scope of supply (Select Only those Detectors, which are part of supply)"	Flame Ionization Detector (FID), Thermal Conductivity Detector (TCD or Hot Wire Detector), Electron Capture Detector (ECD), Nitrogen-Phosphorus Detector (NPD) / Flame Thermionic Detector (FTD), Flame Photometric Detector (FPD), Photo Ionization Detector (PID), Pulsed Flame Photometric Detector (PFPD)
	"Number of FID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of TCD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of ECD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of NPD/FTD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of FPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	"Number of PID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of PFPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	Number of Inlets GC can Simultaneously Support	1, 2, 3
	Number of Detectors (Ports) GC can Simultaneously Support	1, 2, 3, 4
	Number of Capillary Columns provided in GC - inclusive in the scope of Supply	0, 1, 2
	Type of Capillary Column	Polar, Non-Polar, Semi-Polar, NA
	"GC Capillary Column Length (Hint:-Select '0' if Not Applicable)"	0, 30, 60, 100 (meter)
	"Number of Packed Columns provided in GC - inclusive in the scope of Supply (Hint:-Select '0' if Not Applicable)"	0, 1, 2
	"GC Packed Column Length (Hint:-Select '0' if Not Applicable)"	0, 3, 6, 9, 12 (foot)
PERFORMANCE	"GC Detector Minimum Detection Level (MDL) "	< 1.4 pg C/s for FID, < 1.8 pg C/s for FID, < 3.0 pg C/s for FID < 400 pg/ml for TCD, < = 1 microgram/ml for TCD, < = 10 microgram/ml for TCD, < 4.5 fg /sec for ECD, < 5.5 fg /sec for ECD, <= 100 fg/sec for ECD, < 0.1pg N /sec for NPD/FTD, < 0.03 pg P /sec for NPD/FTD, <0.5 pg P/s for NPD/FTD, < 45 fg P /sec for FPD, < 100 fg P /sec for FPD, <2.5 pg S /sec for FPD, <2.5 pg S /sec for FPD, <5 pg S /sec for FPD, <10pg S/s for FPD
COLUMN OVEN (GAS CHROMATOGRAPH - GC)	Maximum Operating Temperature (GC Oven)	upto 400 deg. C, upto 450 deg. C (2) Or higher
	Number of Oven Ramps / Plateaus for Temperature programming	6/7 or more, 9/10 or more, 20/21 or more
	Temperature Ramp Rate - Maximum	120 Deg. C/minute or better, 50 Deg. C/minute or better
PNEUMATIC CONTROL	Number of Injection Ports	1, 2

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Maximum Injection Port heating Temperature	upto 400 deg. C (1), upto 450 deg. C (2) Or higher
	Pressure Range	upto100 psi (1), upto 120 psi (2) Or higher
	Pressure set point Resolution (\pm)(psi)	0.001, 0.01, 0.1 Or lower
	Type of Sample Injection Port	Split, Splitless, split / Splitless, Purged packed Inlet
	Split Ratio	6000:1, 7000:1, 7500:1, NA
AUTO INJECTOR / AUTO SAMPLER	Compatible Liquid Auto Sampler - inclusive in the scope of supply	Yes, No
	Capacity of the Liquid Auto Sampler (Number of Vials)	0, 15, 100, 150 Or higher
	Number of Injection Ports in the Liquid Auto Sampler	0, 1, 2
	Number of Channels in the Liquid Auto Sampler	0, 2
	Minimum Volume of the injected liquid sample	0 - 1 (microliter)
	Maximum Volume of the injected liquid sample	0 - 100 (microliter)
	"Liquid Auto Sampler Features (Select applicable features Only)"	Fully automated Liquid Auto Sampler, Injects with variable speed & varying syringe sizes, Operates through Software (Software inclusive in the scope of supply), NA
HEAD SPACE SAMPLER	Compatible Head Space Sampler- inclusive in the scope of supply	Yes, No
	Capacity of the Head Space Sampler Carousel (number of Vials)	12 or better, NA
	Head Space Sampler Temperature Range	Upto 200 deg. C or better, NA
PYROLYZER SYSTEM	Compatible Pyrolyzer System for pyrolysis of solid and liquid samples- inclusive in the scope of supply	Yes, No
SALIENT FEATURES OF GAS CHROMATOGRAPH (GC)	"Salient Features of GC (Select applicable features Only)"	Automatic leak testing/detection of Pneumatics, Direct setting of split flow rates and ratios, Flow sensor for control & storage of split ratio, Split and Septum Purge gases, Able to direct inject onto packed and wide-bore capillary columns, Oven power turns-off automatically when the lid/door is opened, Over Heat Protection, Display Unit with built-in Key Board, Controlled by External PC System, NA

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SOFTWARE	"Software Features (Select Applicable Features only)"	Compatible with Windows 7 or higher version operating system, Seamless integration & control of all the GC parameters, Interface with Multiple GCs through the comprehensive real-time monitoring and diagnostic software, Real time notification via advisories and indicators, Counters, Chromatographic attributes, Instrument diagnostics, Access to maintenance information, Access to maintenance, run and event logs, Retention Time - Locking module for analyzing target compounds in complex matrices
GENERIC	Display	Through PC (PC is not part of supply), Built-in Display
	Connectivity interface	USB, WIFI, ETHERNET, Bluetooth, RS 232, RJ-45
	Provision to split the sample from single injection to two different detectors and get results for two detectors simultaneously	Yes, No
ACCESSORIES	Gas Supply inclusive in the scope of supply	Nitrogen, Hydrogen, Argon, Helium, Zero Air, Not offered
	Capacity of Gas cylinder	47 liters water capacity, Not offered
	Number of Cylinders of each type supplied - inclusive in the scope of supply	0, 1, 2
	Gas cylinders certified by PESO	Yes, No for Gas Cylinder Not Offered
	Regulator - inclusive in the scope of supply	Dual Stage Regulator, NA
	"Other Accessories inclusive in the scope of supply (Select only those Accessories which are part of supply)"	Gas purification panel for all Gases with moisture trap, Hydrocarbon trap & Oxygen trap, Syringe for manual injection x 05 nos, Auto Sampler Vials x 1000 vials with screw cap, Graphite ferrules x 10 nos, Vessel ferrules x 10 nos, Septa for injector x 100 nos, Inlet liner for splits x 4 nos, Inlet liner for split x 4 nos, O-rings x 10 nos. of each type, Column Cutter x 2 nos, Gas tube Cutter x 2 nos, Tool Kit x 1 no., User / Technical / Maintenance Manual, Standard Samples - NIST / IPL Traceable, NA
ENVIRONMENTAL	Minimum Operating Temperature	0 - 10 (degree Celsius)
	Maximum Operating Temperature	30 - 65 (degree Celsius)
	Operating Humidity (RH) (%) at 40 degree C	Up to 95 (percent)
INSTALLATION & COMMISSIONING / WARRANTY / AVAILABILITY OF SPARES / TRAINING	Warranty on equipment	3.0 (year)
	Number of days onsite training provided	3, 5, 7, 10 (day)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
CERTIFICATION / TEST REPORT	"Availability of UL/CE Certificate as per IEC 61010- 1 latest - Safety requirements for electrical equipment for measurement,control, and laboratory use General requirements"	Yes, No
	Conformance to EMC/EMI as per IEC 61326-1 latest - Electrical equipment for measurement, control and laboratory use	Yes, No
	Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification	Yes, No

Additional Specification Parameters - Gas Chromatography (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
GC (Dual TCD)) : The Technical Specifications strictly must be as per the Attached ATC Document only.	GC (Dual TCD) : The Technical Specifications strictly must be as per the Attached ATC Document only.

* Bidders offering must also comply with the additional specification parameters mentioned above.

परिषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परिषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Rakesh Kumar	851115,HURL BARAUNI PROJECT, BARAUNI URVARAK NAGAR, BEGUSARAI, Begusarai, Bihar, 851115	1	56

Buyer added Bid Specific Additional Scope of Work

क्र.सं./S.No.	Document Title	Description	रिवर्स प्रभार के अनुसार जीएसटी/Applicable i.r.o. Items
---------------	----------------	-------------	--

क्र.सं./S.No.	Document Title	Description	रिवर्स प्रभार के अनुसार जीएसटी/Applicable i.r.o. Items
1	The Specifications strictly must be as per the Attached ATC Document only. View	The Specifications strictly must be as per the Attached ATC Document only.	Gas Chromatography(1), Gas Chromatography(1)

The uploaded document only contains Buyer specific Additional Scope of Work and / or Drawings for the bid items added with due approval of Buyer's competent authority. Buyer has certified that these additional scope and drawings are generalized and would not lead to any restrictive bidding.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

Hindustan Urvarak and Rasayan Limited
Urvarak Nagar
Barauni
Begusarai
Bihar - 851115
.

2. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

3. Generic

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- Copy of PAN Card.
- Copy of GSTIN.
- Copy of Cancelled Cheque.
- Copy of EFT Mandate duly certified by Bank.

4. Generic

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

5. Generic

- The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of buyer.
- The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.
- The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due performance of the Contract and the Sellers obligations there under.

6. Generic

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- i) The Seller fails to comply with any material term of the Contract.
- ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- iv) The Seller becomes bankrupt or goes into liquidation.
- v) The Seller makes a general assignment for the benefit of creditors.
- vi) A receiver is appointed for any substantial property owned by the Seller.
- vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

7. Generic

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

8. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

9. Turnover

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

10. Purchase Preference (Centre)

Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 25% of total value.

11. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

12. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

The Technical specifications Strictly be as per the attached AT C documents only.

13. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---